



Perfect for use in making on-site measurements $\zeta \in$






The Handy Hardness Tester (SONOHARD®) model SH-21(E) differs completely from traditional hardness testers from a viewpoint that instead of measuring the size of the indentation of the test sample using a microscope, it employs a diamond indenter equipped with a vibrating rod that presses on the test surface at a fixed force and then measures its hardness by applying ultrasonic vibrations.

When the vibrating rod is applied to a soft-surfaced test sample with identical qualities and at a fixed force, it makes a deep indentation and gets locked into the groove. Due to this, the resonance frequency increases. Conversely, it does not get locked in when used on hard test samples and the resonance frequency drops. The test sample's hardness can be calculated using the correlation between this deviation and the tested hardness.





Calculation values of SH indentation


Load P= approx. 2kgf (approx. 20N)

Hardness Calculation value, HV Size of indentation (mm)		Calculation value, Depth of indentation (mm)	Conversion value, HRC	
100	0.193	0.028	-	
200	0.136	0.019	15	
300	0.111	0.016	31	
400	0.096	0.014	41	
500	0.086	0.012	49	
600	0.079	0.011	55.5	
700	0.073	0.010	60.5	
800	0.068	0.010	64.5	
900	0.064	0.009	67.5	

Load P= approx. 1kgf (approx. 10N)

Hardness HV	Calculation value, Size of indentation (mm)	Calculation value, Depth of indentation (mm)	Conversion value, HRC
100	0.136	0.019	-
200	0.096	0.014	15
300	0.079	0.011	31
400	0.068	0.010	41
500	0.061	0.009	49
600	0.056	56 0.008 55.	
700	0.051	0.007	60.5
800	0.048	0.007	64.5
900	0.045	0.006	67.5

Probe dimensions



Precautions on measurements

1. The affect of surface roughness


X:Average value		σ:Standard deviation	Measurement frequency per 100	
Hardness	Surface roughness	0.8a	1.6a	3.2a
UDON F	X	31.5	31.7	30.9
HRC31.5	σ	0.4	0.5	0.8
HRC50.8	X	50.5	50.5	50,3
HHC50.8	σ	0.3	0.3	0.6
HRC65.5	X	65.4	65.3	65.1
	a	0.2	0.2	0.4

For items with a roughness of 3.2a or greater, you will need to polish the surface before
making measurements. If decarbonization occurs, make measurements after having removed it.

2. Measurable dimensions (For loading of 2kgt/approx. 20N)

- Size: 15mm wide x 15mm long or greater 2 Thickness t= 7mm or greater
- For items smaller in dimension than those listed above, you will need to use a stand with dimensions of \$50 x 15mm or greater, and make measurements after securing the item to the stand by applying oil and after increasing the virtual mass.
- For loading of 1kgf/ approx. 10N, it is possible to make measurements even at t= 7mm or less.

3. Angles and deviations



4. Reproducibility

When making measurements on a standard test block using a measuring stand

		Standard block	Stand measurement	Hand-held measurement
HV hardness	Average value	HV200	199	201
		500	504	508
		800	806	790

Specifications

Model name	Handy Hardness Tester SONOHARD® SH-21(E) (Motorized/manual switchover type probe)		Display make-up	a. Measured value: 3 digits b. TIMES: 2 digits (measuring frequency) c. MAX value: 3 digits d. MIN value: 3 digits e. σ: 4 digits (standard deviation) f. χ: 4 digits (average value)
Measuring indenter	Diamond indenter for Micro-Vickers (facing-to- surface angle of 136")			
Test load and control no.	1. Approx 2kgf (roughly 20N) SH-21-E2			
CE making complied models are	2. Approx 1kgf (roughly 10N) SH-21-E1			
required to add (CE) with control no.				
Measuring range	1. Rockwell hardness HRC 10~70 2. Vickers hardness HV 50~999 3. Shore hardness HS20.0~99.9 4. Brinell hardness HB 85~550			
Reproducibility	HRC: \pm 1.0HRC, HV: \pm (3%rdg)HV, HS: \pm 1.0HS, HB: \pm (3%rdg)HB	Set-up	a. UPPER (upper limit) b. LOWER (lower limit) c. TIMES (measuring frequency) d. CANCEL	
Applicable test materials	With steel as the principle material, oth materials may also be measured by ca against a standard hardness test block			
Display of measured values	Digital display (LCD, 4 digits) with EL b			
Data memory	2,000 pieces			
Digital display units	1HV, 0.1HRC, 0.1HS, 1HB, 1N/mm2 (to	Alarm	a. Alarm (buzzer sound)	
Allowable operating temperature	0~50°C		Output	a. RS-232C output used for data
Power source	AC adapter(100~240V) or lithium ion i	rechargeable battery		
Continious operating hours	5 hours when featuring a rechargeable BL; 8 hours without BL *BL= backlight			transmission or printing
Dimensions	Display unit: 97(W) x 50(D) x 170(H) m Probe diameter: 40mm, length: 164mm	n	Frequency	1. Motorized / manual switchover type approx 69~71kHz 2. Manual type (special type)
Weight	Display unit: approx. 400g (including be Probe and cable: approx. 400g	attery),		
Carrying case dimensions	400(W) x 220(D) x 140(H) mm			approx 38~40kHz
Standard components	1 display unit, 1 probe (including standa 1 hardness standard test block (arroun 1 probe cable 1.5m, 1 AC adapter 100 1 recharger (MK-8220), 1 lithium ion ba 1 carrying case (MK-9701)	Conversion	HV、HRC、HS、HB、N/mm ² according to JIS B 7731, ASTM E 140, DIN 50150	
Options	Standard hardness test blocks HV600, HS50, HB300 used for scale calibration, Measuring stand for small objects (SH-S02), Probe attachment for pipes materials (UA4-4005), Probe attachment for inner races (UA4-4006), Printer model DPUH245AS with cable, Printer paper (TP-H241L), Stand for the main unit (SH-P03)			

- When using the tester installed in automated machinery, please contact our hardness tester sales department for specifications concerning the testers used for automatic machines.
- In order to make it possible to propose ideas regarding the unique loads, jigs and measuring methods to our customers, we humbly request that you fill out a specification confirmation form when making inquiries concerning usage.
- The SONOHARD SH-21(E) is calibrated using the standard hardness test block produced in compliance with JIS B7730/ ISO 6508-3 and JIS B7735/ ISO 6507-3 by Yamamoto Scientific Tool Laboratory Co., Ltd., Japan, who has the quality management system approved under ISO 9001. The values measured by SH-21(E) are therefore guaranteed by us. (Measured values under the calibration using other makers' test blocks are out of our guarantee.)
- The model name on the catalog is SH-21(E), while it is referred to as SH-21 only in the relevant operation manual, test certificate and ISO certificate, etc.
- A standard export model of SH-21(E) is not CE-Marking complied, but a CE-Marking complied model is also available by factory modification on request and order beforehand.

Please read the user's manual before undertaking operations.
 Specifications may be changed without prior notice due to product revisions.



Tokyo Branch Office (Overseas Sales)

Okaya Bidg., 14-4 Nihonbashi-Kodemma-cho, Chuo-ku, Tokyo 103-0001, Japan

e-mail: tokyo@kac.co.jp Tel. 03-3662-5341 Fax 03-3662-5346

Head Office and Main Plant

3-48 Takahata-cho, Nishinomiya-shi, Hyogo Pref.

663-8202 Japan e-mail: honsha@kac.co.jp Tel. 0798-66-1508 Fax. 0798-65-7025

EC-SH-21 01-A 04.05.5000 (2)

Utilization of the Handy Hardness tester SH-21

Examples of quality control and maintenance usage by measuring handness



Processed goods, Press parts, Metal mold



Tapered parts



Aluminum wheels



 Diagnosing wear and tear using hardness measurements



Drill blade



 Measuring the strength of welding sections (Checking tension strength)

Measuring metal fatigue in steel towers, bridges and reinforcing bars







Measurements with Measuring stand for small objects

